



Certificate of Containment Testing

Containment Testing of Bucket GLBB-bdc-15-100 in a GYROZEN Centrifuge

Report No. 20/110

Report Prepared For: GYROZEN Co., Ltd.

Issue Date: 12 March 2021

Test Summary

Bucket GLBB-bdc-15-100 was tested for containment in a GYROZEN centrifuge, using Annex AA of IEC 61010-2-020:2016 (3rd Ed.). The sealed buckets were shown to contain a spill.

Report Written By

Name: Ms Helen Hookway

Title: Biosafety Scientist

Report Authorised By

Name: Mrs Sara Speight

Title: Senior Biosafety Scientist



Certificate of Containment Testing

Containment Testing of Bucket GLBB-b100-100 in a GYROZEN Centrifuge

Report No. 20/110

Report Prepared For: GYROZEN Co., Ltd.

Issue Date: 12 March 2021

Test Summary

Bucket GLBB-b100-100 was tested for containment in a GYROZEN centrifuge, using Annex AA of IEC 61010-2-020:2016 (3rd Ed.). The sealed buckets were shown to contain a spill.

Report Written By

Name: Ms Helen Hookway

Title: Biosafety Scientist

Report Authorised By

Name: Mrs Sara Speight

Title: Senior Biosafety Scientist